

The 53rd ASAIO Conference, Palmer House Hilton, Chicago, IL, June 7-9, 2007

PLASMA FREE HEMOGLOBIN CLEARANCE BY HEMOCOR HPH®
HEMOCONCENTRATOR AND OTHER COMPETITIVE DEVICES

R Wenthold.¹ I L Thomas IV.² J Simoni,² J F Moeller.² G Simoni.² S London.¹
J Villanueva-Meyer² and C Roongsritong²

¹Minntech Corporation, Minneapolis, MN and ²Texas Tech University Health Sciences Center, School of Medicine, Lubbock, TX

Increased plasma hemoglobin (Hb) is often observed in bypass surgery and ECMO patients. Hb, after exceeding the capacity of its detoxifying system, triggers many pathologic reactions and poses a high risk to coronary artery endothelium. Our previous study showed ineffectiveness of cellulose-based hemodialyzers in Hb removal. The present *ex vivo* study examined free Hb clearance by clinically accepted polysulfone-based hemoconcentrators; Hemocor HPH® 1400 (Minntech Corp.), Capiiox® HCO5 (Terumo Corp., Tokyo, Japan) and Senko CF11 (Senko Medical Trading Co., Tokyo, Japan). All experiments mimicked the 2-hour human bypass surgery procedure, scaled down to a 1.5 L model, using standard blood and ultrafiltrate flow rates with sampling times: 0.0, 0.25, 0.5, 1 and 2 hours. Baseline Hb level in the hemoconcentrator circuit was 0.7 ± 0.03 g/dL. Results indicate that Hemocor HPH® was superior to the other tested devices. While Hemocor HPH® eliminated more than 90% of Hb after 2 hours ($y = 1.374 \times X^{(1.022)}$, $R^2 = 0.986$), the Capiiox® HCO5 and Senko CF11 cleared less than 16% ($p < 0.001$, $y = 0.498 \times X^{(0.875)}$, $R^2 = 0.982$) and 48% ($p < 0.001$, $y = 0.733 \times X^{(0.940)}$, $R^2 = 0.993$), respectively. No tested polysulfone membranes affected the physico-chemical properties of Hb with respect to its tetrameric conformation, heme stability and oxidation, and ferryl-Hb species formation.