

Certificate of Analysis


RAPICIDE[®] PA

Part A

Product Code: ML02-0147
 Lot Number: 00017519
 Expiration Date: 2020/10

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	23.3 %
PAA	4.0 – 6.0 % (w/w)	5.2 %
Acetic Acid	8.0 – 10.0% (w/w)	9.1 %
pH, (as is)	0.5 – 1.1 (as is)	1.0


This lot has been reviewed and meets specified requirements.



 Quality Analyst

30/04/19

 Analysis Date



 Quality Assurance

30-APR-19

 Release Date

Medivators Inc.
 14605 28th Avenue North
 Minneapolis, MN 55447 USA
 Telephone: (763) 553-3300
 Fax (763) 553-3387
 www.medivators.com

RAPICIDE[®] is a registered trademark of Medivators Inc.