



Certificate of Analysis

INTERCEPT™ PLUS

Detergent & Cleaner

Part No. ML02-0159

Lot Number: 00017818

Expiration Date: 2020/05

Property	Test Method	Specification	Analysis
Appearance	Visual	Clear Non-viscous Liquid	Pass
pH (release) at 25°C	20002-154	11.5 – 12.5	12.1
Phosphate Level	20602-105	80.83 – 98.79 g/L	93.61 g/L

This lot has been reviewed and meets specified requirements.

Quality Analyst

18 May 2018
Analysis Date

Quality Assurance

21 May 18
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

INTERCEPT™ is a trademark of Medivators Inc.