

Certificate of Analysis
RAPICIDE[®] PA
Part A

Product Code: **ML02-0147**
Lot Number: 00017918
Expiration Date: 2019/11

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	<u>22.7 %</u>
PAA	4.0 – 6.0 % (w/w)	<u>5.1 %</u>
Acetic Acid	8.0 – 10.0% (w/w)	<u>9.2 %</u>
pH, (as is)	0.5 – 1.1 (as is)	<u>0.9</u>


This lot has been reviewed and meets specified requirements.



Quality Analyst

18 May 2018

Analysis Date



Quality Assurance

18 May 18

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE[®] is a registered trademark of Medivators Inc.