

**Certificate of Analysis**  
**RAPICIDE® PA**  
**Part B**

Product Code: ML02-0148  
Lot Number: 00018018  
Expiration Date: 2019/11

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	786.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	1.0 %w/w
pH	11.4 to 12.4 as is	11.7


This lot has been reviewed and meets specified requirements.



Quality Analyst

18 May 2018

Analysis Date



Quality Assurance

21 May 18

Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Plymouth, Minnesota 55447  
Telephone: (763) 553-3300  
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.