



Certificate of Analysis

INTERCEPT™ PLUS

Detergent & Cleaner

Part No. ML02-0159

Lot Number: 00018019

Expiration Date: 2021/05

Property	Test Method	Specification	Analysis
Appearance	Visual	Clear Non-viscous Liquid	Pass
pH (release) at 25°C	20002-154	11.5 – 12.5	11.9
Phosphate Level	20602-105	80.83 – 98.79 g/L	84.60 g/L

This lot has been reviewed and meets specified requirements.

ADC
Quality Analyst

03/05/19
Analysis Date

[Signature]
Quality Assurance

06-MAR-19
Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

INTERCEPT™ is a trademark of Medivators Inc.