



# Certificate of Analysis

## INTERCEPT™ PLUS

### Detergent & Cleaner

Part No. ML02-0145

Lot Number: 00018119

Expiration Date: 2021/05

Property	Test Method	Specification	Analysis
Appearance	Visual	Clear Non-viscous Liquid	Pass
pH (release) at 25°C	20002-154	11.5 – 12.5	12.0
Phosphate Level	20602-105	80.83 – 98.79 g/L	88.85 g/L

This lot has been reviewed and meets specified requirements.

Quality Analyst

Analysis Date

Quality Assurance

Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)

INTERCEPT™ is a trademark of Medivators Inc.