



Certificate of Analysis

INTERCEPT™ PLUS

Detergent & Cleaner

Part No. ML02-0159

Lot Number: 00022919

Expiration Date: 2021/06

Property	Test Method	Specification	Analysis
Appearance	Visual	Clear Non-viscous Liquid	Pass
pH (release) at 25°C	20002-154	11.5 – 12.5	12.4
Phosphate Level	20602-105	80.83 – 98.79 g/L	86.91 g/L

This lot has been reviewed and meets specified requirements.

Quality Analyst [Signature] Analysis Date 10/06/19

Quality Assurance QP [Signature] Release Date 14/06/2019

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

INTERCEPT™ is a trademark of Medivators Inc.