


Certificate of Analysis
RAPICIDE® PA
Part B

Product Code: ML02-0148
Lot Number: 00023119
Expiration Date: 2020/12

Components	Specification	Analysis
Sodium	743.0 - 909.0 mEq/L	762.0 mEq/L
Benzotriazole	0.9 – 1.1% w/w	0.9 %w/w
pH	11.4 to 12.4 as is	11.8


This lot has been reviewed and meets specified requirements.



Quality Analyst

10/26/19

Analysis Date



Quality Assurance

13-JW-19

Release Date

Medivators Inc.
14605 28th Avenue North
Plymouth, Minnesota 55447
Telephone: (763) 553-3300
Fax (763) 553-3387

RAPICIDE® is a registered trademark of Medivators Inc.