

Certificate of Analysis
RAPICIDE® PA
Part A

Product Code: **ML02-0147**
Lot Number: 00035819
Expiration Date: 2021/03

| Components | Specifications | Analysis |
|-------------------------------|--------------------|---------------|
| H ₂ O ₂ | 20.0 – 24.0% (w/w) | <u>23.0 %</u> |
| PAA | 4.0 – 6.0 % (w/w) | <u>5.1 %</u> |
| Acetic Acid | 8.0 – 10.0% (w/w) | <u>9.2 %</u> |
| pH, (as is) | 0.5 – 1.1 (as is) | <u>0.8</u> |

This lot has been reviewed and meets specified requirements.



Quality Analyst

23-SEP-19

Analysis Date



Quality Assurance

23-SEP-19

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.