

Certificate of Analysis

RAPICIDE® PA

Part A

Product Code: **ML02-0147**

Lot Number: **00039018**

Expiration Date: **2020/05**

Components	Specifications	Analysis
H ₂ O ₂	20.0 – 24.0% (w/w)	23.2 %
PAA	4.0 – 6.0 % (w/w)	5.5 %
Acetic Acid	8.0 – 10.0% (w/w)	8.8 %
pH, (as is)	0.5 – 1.1 (as is)	0.5


This lot has been reviewed and meets specified requirements.



Quality Analyst

30-NOV-18

Analysis Date



Quality Assurance

30-Nov-18

Release Date

Medivators Inc.
14605 28th Avenue North
Minneapolis, MN 55447 USA
Telephone: (763) 553-3300
Fax (763) 553-3387
www.medivators.com

RAPICIDE® is a registered trademark of Medivators Inc.