



**Certificate of Analysis**

**RAPICIDE® PA**

**RTU Part B**

Lot Number: 000418

Expiration Date: 2019/07

Quality Control Report: 180003

Components	Specifications	Analysis
Na+	35.5 – 43.4 mEq/L	<u>35.6 mEq/L</u>
Benzotriazole	0.05% ± 0.005%	<u>0.048 %</u>
pH, (as is)	10.6 – 11.6 (as is)	<u>11.6</u>

This lot has been reviewed and meets specified requirements.

\_\_\_\_\_  
Quality Analyst

03 January 2018

Analysis Date

\_\_\_\_\_  
Quality Assurance

03 JAN 18

Release Date

**Medivators Inc.**  
14605 28th Avenue North  
Minneapolis, MN 55447 USA  
Telephone: (763) 553-3300  
Fax (763) 553-3387  
[www.medivators.com](http://www.medivators.com)